
KERN SET-22 EasyTouch Difference



ET Difference – Difference weighing e.g. growth of cell cultures

Features

- Note: the required basis is SET-01 ET OS (basic package). SET-22 is part of the SET-02 Premium Laboratory package with "recipe function" at a bargain price
- Image: Image: The difference function supports the monitoring and analysis of growth or modification processes, as would occur, for example, in biological series of tests with cell cultures (monitoring). In this function you can define, name and record the related initial weight of as many sample holders as you wish, e.g. petri dishes (with culture mediums) 4. The samples to be monitored are entered in this sample holder and recorded with their initial weights in the function. Using the difference function you can weigh this sample holder at regular intervals again and again. The difference function automatically calculates the weight difference, i.e. the difference in relation to the initial weight (e.g. the growth or other biological, chemical changes). These differences are stored and analysed. Statistical analysis of any differences in a graphical manner and as data to download
- I Graphic step-by-step guide: The illustrated guide takes the user safely through the workflow for difference weighing
- Repeating the weighing sequences: You can repeat the weighing sequence as many times as you wish, i.e. sample holders with cell cultures can be weighed as often as you wish and can therefore be monitored for as long as required

- Is Recalling the relevant sample holder occurs either automatically, in accordance with the defined sequence, or manually using the ID number of the sample holder, which can be scanned using the **barcode** for example. In this way, the **risk of confusion and incorrect weighing** is minimised
- I Central master data memory: These weighing sequences can be stored in the system memory with the number of containers (e.g. petri dishes), container IDs, ID number for the sequence, name of the sequence, batch name, etc. By doing this, this data does not have to be constantly re-entered when repeating the sequence, but can easily be recalled from the memory. The tare values of the petri dishes can also be stored in the master data memory. These are then automatically subtracted from the particular weighing result
- S Efficient weighing and saving of individual results

The sample holders can be stored with an ID number. This can be scanned using a barcode. In this way the correct sample holder is identified. Just weigh, save and you're done

Ň	Difference Statement New Procedure		English- Romer Lister	- 0 X		
		Sumbaryor' Chilan growth				
			_			
	To see the ofference	Officercé / / Sart wegfe Wegterg of same object to define on conditions	۲			
	🖻 Skip Gre segar	22	-			
		Step 2 : Measurement of filled cantainers to dete	more the starting weights			
		Dan't show process screens any more	Proceed			

EasyTouch

New procedure for difference weighing





Recalling the sample holder

	Difference Difference > New Pocedure			English - 🧕 Admin	- D ×
	Please enter the number of samples to be	nessured			8
ଜ	Ranke of angles" 3	Provident spece * Hen weighing			Line of the second
88	Provides incorpore Weighing the growth of hen over a period of time				
	Weighing the growth of hen over a period of the				
0	Skip tare sequence	General General General (8		
	Update object in master memory				
ŀ	Applied master object				
	Value object 0 Halow object 10 Halow object rank Okalem Standard matter 20(1)9989765				
1					

Marking of sample holders


 Master database Oscillare > Edit muster data			English- 👷 Advert Sauter - 🗆			×	
Edit master data							
Component/ Dipol 10* 3456758		Component / Digist name * Chicken			0 runter have 230789098765		
Remaining Chicks weighing		Container weight	9 (11)		Auge factors Difference Procedure		
Only and the grad prov							
Difference Number of samples *	Providera turne						^
3 President discogram	Hen weighing						
Weighing the growth of hen over a period of time							

Central master data memory



Measuring screen

KERN SET-22 EasyTouch Difference





Options

• Save-Server central data memory function for additional storage of all measuring data in a central, local server directory. This is where measuring data is stored from all weighing systems connected using KERN EasyTouch, as well as all installed KERN EasyTouch functions. The advantage of this, particularly for users with several weighing systems, is having all weighing data consolidated in just one database and only having to search for individual measuring data from different balances in one table.

Save-Server data storage is also tamper-proof and cannot be changed, KERN SET-10

• Save-Cloud: Has the same central data memory function as the Save-Data Server for all weighing systems connected to KERN EasyTouch. The difference is that in Save Cloud, the storage location is a KERN Server, which can be accessed over the internet, instead of a server in a local network. Setting up Save-Data Cloud functionality takes place automatically and does not require a network administrator on the user side, KERN SET-101

Technical data

- Licence model: A license can be operated on up to four terminal devices (PCs, laptops, tablets) at the same time and independently
- User: An unlimited number of users can be created in one license
- Balances: You can create and operate as many balances in one licence as you want
- Communication between balance/terminal device: Balances can communicate with the PC, laptop or tablet by serial connection, USB, Bluetooth, Ethernet or WIFI

